

**NORTH DELHI MUNICIPAL CORPORATION
HINDU RAO HOSPITAL; DELHI-7**

NIT No.HRH/ET/2018-19/P-15, dt.01.08.2018.

**TENDER NOTICE
(E-Tendering Mode Only)**

A. E-tenders on dual bid system (technical bid and financial bid.) are invited by the Medical Superintendent, Hindu Rao Hospital on behalf of the Commissioner, North Delhi Municipal Corporation, for supply of **various Laryngeal Endoscopes with Accessories for ENT Department as per specification, Hindu Rao Hospital.**

B. The tender forms with terms and conditions and list of the items, can be viewed and downloaded online on the website www.tenderwizard.com/NORTHDMCETENDER

C. Important Dates and Information.

S.No.	Important Dates	Schedule
1	Start date & time to download the tender	01 st August, 2018 at 18.30 Hours
2	Last date & time to download the tender	27 th August. 2018 at 15.00 Hours
2	Last date & time for online Submission of Technical Bid & Financial Bid	28 th August, 2018 at 15.00 Hours
3	Date & time of opening of Technical Bid	29 th August, 2018 at 15.15 Hours
4	Date & time of opening of Financial Bid	Will be informed later on.
5	Address for Communication	Medical Superintendent, Hindu Rao Hospital, Near Malka Ganj, Delhi-110007. Tel. No. 23905798/23905690 Email –mshrhdmc@gmail.com

D. The following payments to be made online through e-payment mode only:

Mode of Payment towards Tender Document Fee(TDF)& Earnest Money Deposit(EMD) to be paid online through e-Payment mode via :

- i. National Electronic Fund Transfer (NEFT) / Real-Time Gross Settlement RTGS) / Tenderer requires download website generated challan available on e-tender website and make its payment through any of their Bank.
- ii. Internet Payment Gateway (Debit/ Credit Card/Net Banking).

Note: Any Payments made through NEFT/RTGS/OTC will take 24 hours for its reconciliation. Hence the payments through NEFT/RTGS/OTC should be made at least TWO BANK WORKING DAYS in advance before any due date and upload the scanned copy of challans in the e-Tender website as a token of payment.

Cost of tender document	Earnest Money Deposit (EMD)
Rs.500/-	Rs.24725/-

E. The detail of items required with specification is as under :-

S NO	NAME OF THE ITEMS	QTY
1	Laryngoscope – Hopkins Rod 70 Degree	05
2	Laryngoscope – Hopkins Rod 90 Degree	01
3	Flexible Fibreoptic Rhinolaryngoscope	01
4	Xenon Light Source	01
5	Medical Grade Monitor	01
6	Suction Apparatus	02
7	Mobile LED light source (Pen Light)	01

F. TECHNICAL SPECIFICATIONS OF LARYNGEAL ENDOSCOPY EQUIPMENT

REQUIREMENTS FROM THE EQUIPMENT SUPPLIER

- a) The equipment should be of international quality. The manufacturer should be ISO 9001:2008 certified and must have CE approval.
- b) **Three year warranty** from original manufacturer / authorized supplier must be provided for instruments listed below at **Serial No. 1 to 5**.

After warranty expires, the rate of AMC for two years for items at serial no. 1 to 5 should be quoted.
- c) For all other instruments a warranty including repair and/or replacement for at least one year should be provided.
- d) An undertaking from Original Equipment Manufacturer (OEM)/ Authorized Supplier is required that they will be responsible for providing technical support of the equipment including spare parts for ten years.
- e) An undertaking that the firm will provide for preventive maintenance service six monthly during the Warranty period and will also attend to any number of emergency maintenance calls.

S NO	NAME OF INSTRUMENT / EQUIPMENT WITH SPECIFICATIONS	QTY
1	Laryngoscope – Hopkins Rod 70 Degree <ul style="list-style-type: none"> • Outer diameter : 8mm. • High quality stainless steel with good optics. • Autoclavable. • Working length 180 – 190 mm. 	05
2	Laryngoscope – Hopkins Rod 90 Degree <ul style="list-style-type: none"> • Outer diameter : 8mm. • High quality stainless steel with good optics. • Autoclavable. 	01

	<ul style="list-style-type: none"> Working length 180 – 190 mm. 	
3	<p>Flexible Fibreoptic Rhinolaryngoscope</p> <ul style="list-style-type: none"> High resolution fibre optic image guide, 9700 to 10000 pixels. Single fibre diameter of 10 microns, outer diameter of flexible probe between 3.5 to 4.0 mm. High flexibility through increased two-way angulation of distal tip of 140 to 160 degrees. High quality image contrast with increased image diameter. Length 300 mm. Vision direction 0 degrees direct. Field of view 80 to 90 degrees. Working distance 3 mm to infinity. Immersible for liquid disinfection. 	01
4	<p>Xenon Light Source</p> <ul style="list-style-type: none"> Should have minimum 5500 Klx illumination. 250W / 350W Xenon light source 220V. Lamp life min 500 hrs. with easily replaceable bulbs. Compact and sturdy built. Should have intensity control. 	01
5	<p>Medical Grade Monitor</p> <ul style="list-style-type: none"> Reputed make, Medical Grade Monitor 15” or more. 	01
6	<p>Suction Apparatus</p> <ul style="list-style-type: none"> Double bottle Electrical suction apparatus. Adjustable Intensity control. Mobile on castors. 	02
7	<p>Mobile LED light source (Pen Light)</p> <ul style="list-style-type: none"> Rechargeable battery 3 W. Charge time not more than 5 hrs. Lamp life 50,000 hrs. 	01

G. Terms and conditions

- The tender /quotation should be submitted online (technical bid and financial bid separately), along with therequired documents/certificates.
- The rates quoted should be valid for a minimum period of twelve months.
- The tenderer should not be blacklisted or debarred in any govt. institution/department.
- Full specification of each quoted item(s) should be given in technical bid with make and model.
- Supply shall have to be arranged by the firm in 45 days for local item(s) and 90 days for imported item(s).
- All supplies/work will be subject to the approval of board and can be rejected without assigning any reason. Supplier will have to remove rejected supply withinthree days of intimation at own cost.
- No firm will be eligible to withdraw its rates after the submission of the quotation/tender otherwise, EMD will be forfeited.
- Penalty @ 2% per week of the total cost of the order value will be imposed on the firm in case of late supply subject to maximum 6 % for 21 days last form the date of supply order.
- In case the order is placed and the firm is not in position to execute the supplies, the material will be purchased at risk and cost of the firm without any notice.

10. Rates should be clearly written inclusive of all discounts likely to be offered by the firm in the price bid.
11. Rate quoted should be FOR. Supply will be received at the concern store of the hospital.
12. Payment will be made by bill system after receiving the supply and approval.
13. No tender will be accepted without tender cost and earnest money.
14. No previous earnest money will be adjusted.
15. Multiple offers will not be accepted.
16. Percentage of applicable taxes must be clearly mentioned in the technical bid. Also it should be clearly mentioned in the technical bid that the quoted rates are inclusive or exclusive of taxes.
17. The demonstration of the quoted items should be arranged by the bidder as required by the board.
18. Warranty, make and model of the all the quoted items/equipments should be mentioned clearly in the technical bid of tender/quotation.
19. The performance security of 5 % of the cost of the equipment should be deposited by the successful bidder in shape of BG/FDR valid for 60 days beyond the warranty period in favor of the Medical superintendent of the hospital.

20. Documents to be submitted in the technical bid :-

- i. In case of stockiest / distributor quoted the tender, complete chain of the authorization certificate from manufacturer to bidder must be submitted in the technical bid. In case of manufacturer quoted the tender, the document showing the same must be submitted with the technical bid.
 - ii. Proprietorship certificate.
 - iii. GSTIN Registration copy.
 - iv. Forwarding letter on letter head of the firm that the tender conditions are acceptable.
 - v. Undertaking that the firm has not been blacklisted / debarred from anywhere and No CBI/ vigilance enquiry pending against the firm.
 - vi. Copy of PAN.
 - vii. Full specification of quoted item(s) with make and model.
 - viii. The compliance report of the quoted make and model with the NIT specifications, terms and conditions.
 - ix. Any other document such as ISI/ISO/CE/FDA certificate (as per requirement of the NIT specifications, terms and conditions) must also be submitted in the technical bid.
 - x. Undertaking from the manufacturer/Authorized distributor regarding commitment to supply / service of the ordered item/equipment within the stipulated time.
 - xi. List of spares and consumables items should be mentioned in Technical Bid but the rates of AMC and the spares/consumable items should be quoted separately in the Price Bid.
 - xii. Undertaking that the quoted rates are reasonable and justified and same as being charged by them from other Govt. institutions / hospitals.
 - xiii. The detail of service centers in the NCR with complete address, phone no, e-mail ID must be submitted with the technical bid.
 - xiv. Declaration regarding GST, whether quoted rates are inclusive or exclusive of GST. Percentage of applicable GST must be clearly mentioned with HSN code.
21. The quoted firm is responsible for the authenticity of all submitted documents.
 22. No conditional tender/ quotation will be accepted.
 23. The Tenderer should deliver and install the equipment in the concerned department of the Hospital, give the demonstration, and train the staff of the department free of cost.

24. The Medical Superintendent, HRH reserves the right to reject any or all tenders without assigning any reasons.
25. In case the tender/bidder does not submit the required documents, tender cost & EMD, may result in to rejection of the tender/quotation.
26. Any contravention to the above terms and condition may result in to forfeiting of earnest money and rejection of the tender/quotation.
27. The bidder has to fill their rates for the equipments, Rates of AMC and Rates of spares / consumables in the price bid provided with the tender in excel format.

Medical Superintendent.