

**NORTH DELHI MUNICIPAL CORPORATION
HINDU RAO HOSPITAL; DELHI-7**

NIT No.HRH/ET/2018-19/P-22, dt.10.08.2018.

**TENDER NOTICE
(E-Tendering Mode Only)**

A. E-tenders on dual bid system (technical bid and financial bid.) are invited by the Medical Superintendent, Hindu Rao Hospital on behalf of the Commissioner, North Delhi Municipal Corporation, for supply of **Bipolar Resectoscope with Electro Surgical Unit as per specifications for Urology Department, Hindu Rao Hospital.**

B. The tender forms with terms and conditions and list of the items, can be viewed and downloaded online on the website www.tenderwizard.com/NORTHDMCETENDER

C. Important Dates and Information.

S.No.	Important Dates	Schedule
1	Start date & time to download the tender	10 th August, 2018 at 18.30 Hours
2	Last date & time to download the tender	04 th September,2018 at 15.00 Hours
2	Last date & time for online Submission of Technical Bid & Financial Bid	05 th September,2018 at 15.00 Hours
3	Date & time of opening of Technical Bid	06 th September, 2018 at 15.15 Hours
4	Date & time of opening of Financial Bid	Will be informed later on.
5	Address for Communication	Medical Superintendent, Hindu Rao Hospital, Near Malka Ganj, Delhi-110007. Tel. No. 23905798/23905690 Email –mshrhdmc@gmail.com

D. The following payments to be made online through e-payment mode only:

Mode of Payment towards Tender Document Fee(TDF)& Earnest Money Deposit(EMD) to be paid online through e-Payment mode via :

- i. National Electronic Fund Transfer (NEFT) / Real-Time Gross Settlement RTGS) / Tenderer requires download website generated challan available on e-tender website and make its payment through any of their Bank.
- ii. Internet Payment Gateway (Debit/ Credit Card/Net Banking).

Note: Any Payments made through NEFT/RTGS/OTC will take 24 hours for its reconciliation. Hence the payments through NEFT/RTGS/OTC should be made at least TWO BANK WORKING DAYS in advance before any due date and upload the scanned copy of challans in the e-Tender website as a token of payment.

Cost of tender document	Earnest Money Deposit (EMD)
Rs.500/-	Rs.37500/-

E. The detail of items required with specification is as under :-

S NO	NAME OF THE ITEMS	QTY
1	Purchase of Bipolar Resectoscope with Electro Surgical Unit for Urology Department as per specification	01 no.

F. Technical SPECIFICATION FOR BIPOLAR RESECTOSCOPE WITH ELCTRIO SURGICAL UNIT

BIOPOLAR RESECTOSCOPE SET

1. Resection in saline solution.
2. System should function as per principle of true bipolar application and current should return through the same instrument as return path and should avoid body or other instrument as a medium, i.e. the active and return electrode should be on the loop itself which should be completely isolated from telescope, sheath and working assembly thereby sending the current back to machine without using body or resectoscope sheath as a medium.
3. The set should be supplied with one 4 mm 30-degree telescope, autoclavable.
4. Specially designed Bipolar cutting loops, 24 Fr, should be supplied –x 6 No. (Compatible with 12 or 30 degree telescope).
5. Specially designed Collin’s knives Bipolar should be supplied –x 6 No. (compatible with 12 or 30 degree telescope).
6. Specially designed loops for bipolar Resection of Bladder tumours should be supplied to have maximum safety for posterior wall tumours –x 6 No. (Compatible with 12 or 30 degree telescope).
7. Special electrode for Vaporisation of prostate x 06 No.
8. Bipolar Needle electrode x 06 No.
9. Roller Ball Electrode x 06 No.
10. Mushroom loop x 06 No.
11. 24/26 Fr. Inner/outer resectoscope sheaths for continuous flow.
12. It should be supplied with Visual Obturator compatible with resectoscope sheath.
13. It should be supplied with Bipolar, working element, passive cutting by means of spring, movable thumb support, in rest position the electrode is inside the sheath x 1 No.
14. Compatible HF cord x 02 Nos.
15. Ellik evacuators x 01 No.
16. Cystoscope sheath 19-20 Fr. x 01 No.
17. Cystoscope sheath 21-22 Fr. x 01 No.
18. Compatible Double channel bridge x 01 No.

ELECTRO SURGICAL UNIT

1. Generator should have the following outputs: Monopolar, Standard Bipolar, and Saline Bipolar for TURP. System should be upgradable onsite to advance bipolar facility for vessel sealing.
2. For Bipolar saline prostate application: Generator should have saline detection facility.
3. Degree of coagulation can be reselected in several steps.
4. Auto start function: manual adjustment of operating time limit for bipolar coagulation.
5. Voltage regulated cutting.
6. Permanent monitoring of the neutral electrode during monopolar use.
7. The system should be compatible to 1 or 2 pedal foot switches at any point can be used.

8. Should have different precise monopolar and bipolar modes, so that the same ESU can be used for all applications.
9. Generator should have auto start and auto stop function for bipolar modes.
10. It should have 11 cutting options in Unipolar, 10 coagulation for unipolar and 8 bipolar mode.
11. Should have a technology for uniform and smooth cutting even with varying tissue impedance.
12. The Bipolar Saline HF unit should be supplied with a separate Resection Box Module which can be used for performing Bipolar Saline applications.
13. Bipolar application with NaCl irrigation solution.
14. Should have technology for immediate start of cutting.
15. Connecting sockets for unipolar and bipolar application can be selected according to individual requirements.
16. Activation of HF functions possible via footswitch or manual control switch.
17. Convenient use via Touch Screen with external buttons.
18. Switchover function enables switching between two modes within a user.
19. Interdisciplinary use.
20. Spray coagulation.
21. Should have facility to monitor the return electrode contact and should give an alarm.
22. Should have facility to connect separate monopolar and bipolar foot pedal.
23. Should have provision for automatic mist and smoke evacuation facility.
24. The unit should have the generator for more than 300 Watts of power.
25. LF/HF leakage current monitoring/ alarms should be inbuilt feature of unit.
26. The HF unit and entire resectoscope set and its accessories, should be from same manufacturer.
27. All above Equipment and accessories should be certified.

General Terms & Conditions

1. All Equipments Should be US FDA/European CE
2. All equipments' should be from the same manufacturer.
3. The principal Company should have their own Service Centre in Delhi/NCR.
4. Comprehensive Warranty of all equipment's should be of 05 years and 05 years CMC.

G. General Terms and conditions

1. The tender /quotation should be submitted online (technical bid and financial bid separately), along with the required documents/certificates.
2. The rates quoted should be valid for a minimum period of twelve months.
3. The tenderer should not be blacklisted or debarred in any govt. institution/department.
4. Full specification of each quoted item(s) should be given in technical bid with make and model.
5. Supply shall have to be arranged by the firm in 45 days for local item(s) and 90 days for imported item(s).
6. All supplies/work will be subject to the approval of board and can be rejected without assigning any reason. Supplier will have to remove rejected supply within three days of intimation at own cost.
7. No firm will be eligible to withdraw its rates after the submission of the quotation/tender otherwise, EMD will be forfeited.
8. Penalty @ 2% per week of the total cost of the order value will be imposed on the firm in case of late supply subject to maximum 6 % for 21 days last from the date of supply order.
9. In case the order is placed and the firm is not in position to execute the supplies, the material will be purchased at risk and cost of the firm without any notice.

10. Rates should be clearly written inclusive of all discounts likely to be offered by the firm in the price bid.
11. Rate quoted should be FOR. Supply will be received at the concern store of the hospital.
12. Payment will be made by bill system after receiving the supply and approval.
13. No tender will be accepted without tender cost and earnest money.
14. No previous earnest money will be adjusted.
15. Multiple offers will not be accepted.
16. Percentage of applicable taxes must be clearly mentioned in the technical bid. Also it should be clearly mentioned in the technical bid that the quoted rates are inclusive or exclusive of taxes.
17. The demonstration of the quoted items should be arranged by the bidder as required by the board.
18. Warranty, make and model of the all the quoted items/equipments should be mentioned clearly in the technical bid of tender/quotation.
19. The performance security of 5 % of the cost of the equipment should be deposited by the successful bidder in shape of BG/FDR valid for 60 days beyond the warranty period in favor of the Medical superintendent of the hospital.

20. Documents to be submitted in the technical bid :-

- i. In case of stockiest / distributor quoted the tender, complete chain of the authorization certificate from manufacturer to bidder must be submitted in the technical bid. In case of manufacturer quoted the tender, the document showing the same must be submitted with the technical bid.
 - ii. Proprietorship certificate.
 - iii. GSTIN Registration copy.
 - iv. Forwarding letter on letter head of the firm that the tender conditions are acceptable.
 - v. Undertaking that the firm has not been blacklisted / debarred from anywhere and No CBI/ vigilance enquiry pending against the firm.
 - vi. Copy of PAN.
 - vii. Full specification of quoted item(s) with make and model.
 - viii. The compliance report of the quoted make and model with the NIT specifications, terms and conditions.
 - ix. Any other document such as ISI/ISO/CE/FDA certificate (as per requirement of the NIT specifications, terms and conditions) must also be submitted in the technical bid.
 - x. Undertaking from the manufacturer/Authorized distributor regarding commitment to supply / service of the ordered item/equipment within the stipulated time.
 - xi. List of consumable items (Not cover under warranty & CMC) should be mentioned in Technical Bid but the rates of CMC and the consumable items should be quoted separately in the Price Bid.
 - xii. Undertaking that the quoted rates are reasonable and justified and same as being charged by them from other Govt. institutions / hospitals.
 - xiii. The detail of service centers in the NCR with complete address, phone no, e-mail ID must be submitted with the technical bid.
 - xiv. Declaration regarding GST, whether quoted rates are inclusive or exclusive of GST. Percentage of applicable GST must be clearly mentioned with HSN code.
21. The quoted firm is responsible for the authenticity of all submitted documents.
 22. No conditional tender/ quotation will be accepted.
 23. The Tenderer should deliver and install the equipment in the concerned department of the hospital, give the demonstration, and train the staff of the department free of cost.
 24. The Medical Superintendent, HRH reserves the right to reject any or all tenders without

assigning any reasons.

25. In case the tender/bidder does not submit the required documents, tender cost &EMD, may result in to rejection of the tender/quotation.
26. Any contravention to the above terms and condition, may result in to fore fitting of earnest money and rejection of the tender/quotation.
27. The bidder has to fill their rates for the equipment, CMC and consumables, not covered under warranty and CMC, in the price bid provided with the tender in excel format.

Medical Superintendent.